

medax

LUX/LUX 2

SEMI-AUTOMATIC SPRING LOADED BIOPSY SYSTEM
WITH DETACHABLE CANNULA



Semi-automatic spring loaded biopsy system
with adjustable penetration depth and detachable cannula.

- Detachable cannula for multiple procedures.
- 22 mm (Lux) or adjustable 15/22 mm (Lux 2) penetration depth.
- Echogenic treatment for ultrasound guided procedures.
- Centimeter markings and optional sliding stopper.
- Also available with Trocar tip and Sword tip.
- Sterilized by ETO, shelf life 5 years.

medax
medical devices

Your partner in Biopsy and Special Needles

LUX/LUX2

SEMI-AUTOMATIC SPRING LOADED BIOPSY SYSTEM WITH DETACHABLE CANNULA

GAUGE	COLOUR	LENGTH (MM)							
		080	100	120	160	180	200	250	300
14	● Green	14 080	14 100	14 120	14 160	14 180	14 200	14 250	14 300
16	○ White	16 080	16 100	16 120	16 160	16 180	16 200	16 250	16 300
18	● Pink	18 080	18 100	18 120	18 160	18 180	18 200	18 250	18 300
20	● Yellow	20 080	20 100	20 120	20 160	20 180	20 200	20 250	20 300

Different sizes on demand

Ordering information:

LUX NORMAL TIP	L X	GAUGE	LENGTH	0 0
LUX 2 NORMAL TIP	L T	GAUGE	LENGTH	0 0
LUX TROCAR TIP	L X	GAUGE	LENGTH	0 1
LUX 2 TROCAR TIP	L T	GAUGE	LENGTH	0 1
LUX SWORD TIP	L X	GAUGE	LENGTH	0 2
LUX 2 SWORD TIP	L T	GAUGE	LENGTH	0 2

Single box: 10 PCS.

Medax Srl Unipersonale

Headquarters: Via S. Pertini, 4 • 41039 • San Possidonio (MO) • Italy
 Company direct No. : +39 0535 1812757 • Fax No : +39 0535 1812744
 email: customercare@medax.it • PEC: medax@legalmail.it • www.medax.it

Registered Office: Via R. Piva, 1/A • 46025 • Poggio Rusco (MN) • Italy
 Vat N. /Fiscal Code N. Iscriz. Reg. Impr.: MN 02669860369 • N. REA: MN 233527, MO 403036 • Capitale Sociale Euro 100.011,00 i.v.



Medax's quality management system is certified to ISO 13485: 2016 standards.
 In accordance with the requirements of the medical device 93/42/EEC directive and its relevant updates.
 All products undergo intensive clinical testing and are fully EC and FDA approved.