

**medax**

# MEDBONE

BONE MARROW BIOPSY SYSTEM



FISH MOUTH TIP

SPECIAL DIAMOND TIP

Bone marrow biopsy needle to perform bone marrow biopsy “MB” in the iliac crest.

- Coloured-code ergonomic handle.
- Available in different tip configurations: Diamond, Aspiration and Fish Mouth.
- Luer Lock connection for syringe.
- Remover guide provided with indicators to facilitate the sample expulsion and to check its length during the procedure.
- Provided with a special cap for a safe removal of the sample.
- Sterilized by ETO, shelf life 5 years.

**medax**  
medical devices

Your partner in Biopsy and Special Needles

**MEDaX**

# MEDBONE

## BONE MARROW BIOPSY SYSTEM

GAUGE	COLOUR	LENGTH (MM)		
8	● Dark Green	08 070	08 100	08 150
9	● Pink	09 070	09 100	09 150
11	○ White	11 070	11 100	11 150
13	● Light blue	13 070	13 100	13 150

Different sizes on demand

### Ordering information:

MEDBONE DIAMOND TIP

M	B	GAUGE	LENGTH	0	0
		<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>

MEDBONE ASPIRATION TIP

M	B	GAUGE	LENGTH	0	1
		<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>

MEDBONE FISH MOUTH TIP

M	B	GAUGE	LENGTH	0	2
		<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>

Single box: 10 PCS.

#### Medax Srl Unipersonale

**Headquarters:** Via S. Pertini, 4 • 41039 • San Possidonio (MO) • Italy  
Company direct No. : +39 0535 1812757 • Fax No : +39 0535 1812744  
email: customercare@medax.it • PEC: medax@legalmail.it • www.medax.it

**Registered Office:** Via R. Piva, 1/A • 46025 • Poggio Rusco (MN) • Italy  
Vat N./Fiscal Code N. Iscriz. Reg. Impr.: MN 02669860369 • N. REA: MN 233527, MO 403036 • Capitale Sociale Euro 100.011,00 i.v.



Medax's quality management system is certified to ISO 13485:2016 standards.  
In accordance with the requirements of the medical device 93/42/EEC directive and its relevant updates.  
All products undergo intensive clinical testing and are fully EC and FDA approved.